



*International Symposium for Thin Layer Chromatography*

**HPTLC 2006, BERLIN - :- 9<sup>th</sup>-11<sup>th</sup> October 2006**

tc { vguv

**Radioactivity detection in HPTLC**

raytest Isotopenmessgeraete GmbH

Benzstr.4


D-75334 Straubenhardt

Tel. +49(0)7082/9255-0

Fax. +49(0)7082/20813

www.raytest.com

tc { vguv"p wengct kpuvtwo gpvu



tc {vguv

integrated Radio TLC detection systems  
based on:

**25 years experience in Radio TLC detection**  
**actual international regulations**  
**own hardware development**  
**own software development**  
**own production & support**  
**job-training on-site**

**Correct instrument selection**



## Target (wish):

- **Reliable method**
- **Simple to perform**
- **Cheap chromatography equipment**
- **Highest resolution**
- **High homogeneity**
- **Highest sensitivity**
- **Highest dynamic range & linearity**
- **Repeatable measurement**
- **Fast measurements**
- **One for everything**
- **Compliant to related regulations (GLP/GMP/21CFRpart11)**





The golden key

=

perfect solution for  
everything

or in german the

“Eierlegende Wollmilchsau”



Required resolution

Required LOD

Required dynamic

money



Used isotopes

Related regulations

Sample size

Number of samples

homogeneity

tc {vguv"P wengct kpuvtwo gpvu



money

Used isotopes

Sample size

Number of samples  
/ speed


Required resolution


Required LOD

Required dynamic

Related regulations











miniGita Star beta



Gita Star gamma



miniGita Star gamma



Gita Star beta



Marita Star beta

money





Phosphor imager




Rita Star beta

tc {vguv"P wengct kpuvtwo gpvu



Master resolution 25 $\mu$ m








- position sensitive proportional counter
- fast measurement along the total trace
- high sensitive for beta- or beta+ radiation (3H 100dpm/peak 14C 10dpm/peak)
- low sensitive for gamma radiation
- Requires P10 gas
- Resolution 3H 0.5mm 14C 1mm
- Ultra high dynamic range 1 : 1 000 000
- Low BKG 0,1 cpm / mm
- Live display


Marita Star beta






Rita Star










Scintillation probe  
 BGO crystal  
 V-shape  
 high sensitivity (gamma)  
 high resolution  
 Tungsten shielding






### miniGita / Gita gamma resolution 25 $\mu$ m







**Scintillation probe**  
**NaJ(Tl)**  
**1x1“**  
**15 mm lead shielding**



TLC



**Scintillation probe**  
**NaJ (Tl)-T-shape**  
**3 mm thick**  
**20 mm wide**  
**50 mm high**



TLC

tc {vguv"p wengct kpuvtwo gpvu



# Regulations

- GLP
- GMP
- Pharmacopoeia
- 21CFRpart11





### Method Editor

Method: Jim

Counting | Solvent | Comment | Evaluation | GLP

Counting time: 1 min

max. Counts: 0

Scan range [mm]: 0 .. 100

Energy range [keV]: 30 .. 750

Binning: 2

OK Cancel Help

### Mini GINA Method Editor

Method: Jim

Counting | Solvent | Comment | Evaluation | GLP

GLP compatible

Changeable parameters

Counting       Comment

Solvent         Evaluation

OK Cancel Help


### Report

Trace: TLC

Print with graphic Preview Print Qk

Sample description Integration

Substance	R/F	%Total	Type	Area	%Area
		%		CNT*mm	%
No name 001	0.060	8.99	DD	1293.000	9.36
No name 002	0.095	7.32	DD	1053.000	7.62
No name 003	0.124	6.14	DD	883.667	6.40
No name 004	0.148	5.84	DD	840.333	6.08
No name 005	0.165	5.10	DD	734.333	5.32
No name 006	0.177	5.64	DD	811.333	5.87
No name 007	0.186	4.18	DD	601.000	4.35
Substance-008	0.311	3.67	DD	527.333	3.82
No name 008	0.325	4.96	DD	714.000	5.17
No name 009	0.348	2.29	DD	329.000	2.38
No name 010	0.359	4.62	DD	664.333	4.81
No name 011	0.389	4.88	DD	701.667	5.08
No name 012	0.413	6.19	DD	891.000	6.45
No name 013	0.430	5.89	DD	846.667	6.13
No name 014	0.439	4.72	DD	679.333	4.92
No name 015	0.452	7.27	DD	1045.667	7.57
Substance-017	0.471	0.33	DD	47.667	0.35
No name 016	0.589	4.66	DD	671.000	4.86
No name 019	0.940	3.32	DD	478.000	3.46



### Report

Trace: TLC

Print with graphic Preview Print Qk

Sample description Integration

Measurement: Finger\_UL600.rta, started: 01.01.2000

Method: 1 from: 01.01.2000

Origin: 16 mm Front 202 mm

Resolution-Test mit Amersham-Platte "Finger"

HV= 1640V Amp. max(high)

LL= 100 UL= var.

terminated after 00:05:00

There is no golden key (one for everything)



But there is a best solution for a specific application



Thank you for your attention!

